

Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	22510-1K	2026-03-05
Low	22510-11	2026-03-05
Normal	22510-12	2026-03-05
High	22510-13	2026-03-09

Open vial stability 14 days ¹

SWELAB

CONTROL		L	
LOT		22510-11	
Swelab Alfa/Alfa Plus		Mean ²	± Range ³
RBC	10 ¹² /L	2.25	0.12
MCV	fL	68.7	5.0
RDW	%	13.7	4.0
Hct	%	15.5	2.5
Plt	10 ⁹ /L	83	15
MPV	fL	12.4	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.7	0.3
MCH	pg	25.3	1.5
MCHC	g/dL	36.8	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	47.6	7.0
Mid	%	3.8	5.0
Gran	%	48.6	7.0

CONTROL		N	
LOT		22510-12	
		Mean ²	± Range ³
		4.17	0.18
		83.0	5.0
		12.6	4.0
		34.6	3.0
		246	30
		12.7	1.5
		8.5	0.6
		12.0	0.4
		28.8	1.5
		34.7	3.0
		4.3	0.6
		0.3	0.5
		3.9	1.0
		50.3	5.0
		3.7	5.0
		46.0	5.0

CONTROL		H	
LOT		22510-13	
		Mean ²	± Range ³
		5.03	0.22
		93.6	5.0
		11.9	4.0
		47.1	3.5
		544	60
		12.4	1.5
		20.2	1.8
		16.1	0.5
		32.0	1.5
		34.2	3.0
		12.8	1.5
		0.7	1.0
		6.7	2.5
		63.7	5.0
		3.3	5.0
		33.0	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αεροστεγήσση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelverdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozimë
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Rasjon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Orseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
 Lot: 22510-11
 Prod.: Con-Diff Low
 ID: 2251011+
 Exp.: 5-Mar-2026

Instr.: Swelab Alfa
 Lot: 22510-12
 Prod.: Con-Diff Norm
 ID: 2251012+
 Exp.: 5-Mar-2026

Instr.: Swelab Alfa
 Lot: 22510-13
 Prod.: Con-Diff High
 ID: 2251013+
 Exp.: 9-Mar-2026

