

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹

Control	Lot	
Tri-Level	22607-1K	2026-11-30
Low	22607-11	2026-11-30
Normal	22607-12	2026-11-30
High	22607-13	2026-11-30

SWELAB

CONTROL		L	
LOT		22607-11	
Swelab Alfa/Alfa Plus		Mean ²	± Range ³
RBC	10 ¹² /L	2.32	0.12
MCV	fL	70.2	5.0
RDW	%	13.9	4.0
Hct	%	16.3	2.5
Plt	10 ⁹ /L	78	15
MPV	fL	11.8	1.5
WBC	10 ⁹ /L	3.3	0.3
Hgb	g/dL	5.9	0.3
MCH	pg	25.4	1.5
MCHC	g/dL	36.2	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.6	0.8
Lymph	%	47.9	7.0
Mid	%	4.1	5.0
Gran	%	48.0	7.0

CONTROL		N	
LOT		22607-12	
		Mean ²	± Range ³
		4.18	0.18
		83.6	5.0
		12.4	4.0
		34.9	3.0
		242	30
		11.8	1.5
		8.5	0.6
		12.0	0.4
		28.7	1.5
		34.4	3.0
		4.3	0.6
		0.4	0.5
		3.8	1.0
		50.9	5.0
		4.4	5.0
		44.7	5.0

CONTROL		H	
LOT		22607-13	
		Mean ²	± Range ³
		5.05	0.22
		94.8	5.0
		11.8	4.0
		47.9	3.5
		557	60
		12.0	1.5
		20.9	1.8
		16.1	0.5
		31.9	1.5
		33.6	3.0
		13.5	1.5
		0.8	1.0
		6.6	2.5
		64.7	5.0
		3.8	5.0
		31.5	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilitet otvorene bočice 14 dni (hr)
cs	Stabilita otevřených ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud vialis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διάρθρωση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiolek 14 dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvoreneje injekčneje liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

Sredno
Sredstvo
Průměr
Middelverdi
Keskmine
Moyenne
Mittelwert
Mέση
Media
Nožmē
Vidurkis
Middelverdi
Średnia
Média
Medie
Среднее
Srednja vrednost
Priemer
Povprečno
Media
Medelvärde
Ortalama

3

Обхват
Raspon
Rozsah
Område
Vahemik
Intervalle
Bereich
Αναμενόμενο εύρος α
Intervallo
Diapazons
Intervalas
Område
Zakres
Intervalo
Interval
диапазон
Opseg
Rozsah
Območje
Intervalo
Intervall
Aralığı



www.boule.com



Boule Medical AB
Fagerstagatan 7
SE-163 53 Spånga,
Sweden

25121-8

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa

Lot: 22607-11

Prod.: Con-Diff Low

ID: 2260711+

Exp.: 30-Nov-2026

Instr.: Swelab Alfa

Lot: 22607-12

Prod.: Con-Diff Norm

ID: 2260712+

Exp.: 30-Nov-2026

Instr.: Swelab Alfa

Lot: 22607-13

Prod.: Con-Diff High

ID: 2260713+

Exp.: 30-Nov-2026

